

Cholestech LDX

Level 1 and Level 2 ALT Control Material Assay Sheet

Expiration Date: June 10, 2004

Lot# 3344

Level 1 (White Cap)

Lot No. 33440503

Level 2 (Black Cap)

Lot No. 33440504

Constituent

ALT

Expected Range

24-45 U/L

Expected Range

112-205 U/L

STORAGE:

Do Not Freeze. Store refrigerated at 2° – 6° C (36° – 43° F).

Open vials are stable for 30 days refrigerated at 2° – 6° C (36° – 43° F).

SAMPLE SETTING:

Set the sample type to SAMPLE = "Whole B." in the Cholestech LDX Configuration menu.

CHOLESTECH

3347 Investment Boulevard, Hayward, California 94545-3808

1-800-733-0404

AS3344

FRM11807J Rev C

Issue Date: 11/24/03

*****Sample*****

NOT INTENDED FOR USE