

Sample Expected Ranges



REF 88770

Kit **LOT**

250037V



YYYY/MM/DD

2012/01/20

Cholestech LDX®
Calibration Verification

CAL **VER** **L1** **L2** **L3** **L4**

Alere San Diego, Inc.
San Diego, CA 92121 USA
www.alere.com

Expected Values / DA-Forventede værdier /
DE-Erwartungswerte / EL-Αναμενόμενες
τιμές / ES-Valores esperados / FR-Valeurs
attendues / IT-Valori previsti /
NO-Forventede verdier / PT-Valores
esperados / SK-Očakávané Hodnoty /
SV-Förväntade värden

PN: 28192 Rev. A

CAL VER		TC	HDL	TRG	GLU
L1	mg/dL	111-157	22-35	108-159	69-102
	mmol/L	2.87-4.06	0.57-0.91	1.22-1.80	3.83-5.66
L2	mg/dL	159-224	31-48	149-219	164-242
	mmol/L	4.11-5.79	0.80-1.24	1.68-2.47	9.10-13.4
L3	mg/dL	243-342	35-68	225-331	248-365
	mmol/L	6.28-8.85	0.91-1.76	2.54-3.74	13.8-20.3
L4	mg/dL	321-451	61-96	340-501	347-500
	mmol/L	8.30-11.7	1.58-2.48	3.84-5.66	19.3-27.8

Sample Expected Ranges