

Cholestech LDX

Level 1 and Level 2 Lipid Control Material

Assay Sheet

Expiration Date: December 17, 2004

Lot# 3351

Level 1 (Red Cap)

Lot No. 33510505

Level 2 (Black Cap)

Lot No. 33510506

<u>Constituent</u>	<u>Expected Range</u>		<u>Expected Range</u>	
	Mg/dL	mM/L	Mg/dL	mM/L
Total Cholesterol	137-193	3.54-4.99	216-304	5.59-7.86
HDL Cholesterol	25-39	0.65-1.01	50-78	1.29-2.02
Triglycerides	111-163	1.25-1.84	222-327	2.51-3.69
Glucose	105-154	5.83-8.55	210-309	11.66-17.15

STORAGE:

Do Not Freeze. Store refrigerated at 2° – 8° C (36° – 46° F).

Open vials are stable for 30 days refrigerated at 2° – 8° C (36° – 46° F).

SAMPLE SETTING:

Set the sample type to SAMPLE = "Whole B." in the Cholestech LDX Configuration menu.

CHOLESTECH

3347 Investment Boulevard, Hayward, California 94545-3808

1-800-733-0404

AS3351

FRM11807M Rev B

Issue Date: 12/23/03

*****Sample*****

NOT INTENDED FOR USE