

Alere HemoPoint® H2 System



Rapid Hemoglobin Results

The Alere HemoPoint® H2 Meter provides a fast, reliable measurement of an individual's hemoglobin level. Hemoglobin testing plays an important role in managing anemia and other red blood cell disorders.



CLIA
WAIVED



Call 1.877.441.7440 or visit alere.com today.

Rapid Hemoglobin and Hematocrit Results

Fast. Using a single drop of blood, the Alere HemoPoint® H2 System offers accurate results for hemoglobin and an estimated hematocrit level*—in less than a minute. The easy-to-read touch screen displays the current time, battery status and allows for easy operation.

Advanced. Utilizing proven microcuvette technology, the Alere HemoPoint® H2 System features a proprietary “soft load” cuvette holder to minimize blood contamination of the meter. The hemoglobin result is flagged if it is outside the user-defined limits.

Easy to Use. The Alere HemoPoint® H2 System’s user-friendly sampling technique minimizes training time and its rechargeable battery makes it completely mobile. Recall the last 4,000 test results with the touch of a button.

Alere HemoPoint® H2 System Products

Alere HemoPoint® H2 Meter

Catalog No. 55118

Includes HemoPoint® H2 Photometer, user guide, quick reference guide, optics cleaner, control cuvette, and power adaptor

Alere™ Hemoglobin Controls

Catalog No. 88772

Includes set of two dropper bottles:
1 x 1.5 mL-low, 1 x 1.5 mL-high

Alere HemoPoint® H2 Microcuvettes

Catalog No. 80549

100 Test Kit - Includes two containers of 50 cuvettes each

Alere HemoPoint® H2 Optics Cleaner

Catalog No. 52219

Includes 5 cleaning units

Product Specifications

Methodology	Optical absorption photometry
Principle	Azidemethemoglobin
Calibration	Factory calibrated
Sample Type	Capillary, venous, or arterial blood
Sample Size	8 µL
Hematocrit Range	36–54% (calculated)
Measuring Range	0–25.6 g/dL
Within Run Precision CV	0.6–0.7%
Total Precision CV	1.2–1.5%
Measuring Time	10–60s



Call 1.877.441.7440 or visit alere.com today.

*Hematocrit result is calculated when hemoglobin result is within 12–18 g/dL
© 2012 Alere. All rights reserved. The Alere Logo and Alere are trademarks of the Alere group of companies. HemoPoint is a trademark of Stanbio Laboratory L.P. under license. PN 1000250-02 9/12